

**EU DECLARATION OF CONFORMITY**

Name and address of the manufacturer: / **Shenzhen Denco Medical Co., Ltd**  
Address: Room 3108, Block 6, Tian'an Cloud Park , Bantian street, Longgang District, Shenzhen, 518129 , Guangdong P.R.China  
Factory Address: Room 101,201,301B,401&801, Block 5, Renluo 6 Road No. 6, Shuikou Street, Huicheng District, Huizhou,516000 Guangdong P.R.China  
(SRN: CN-MF-000014271)

**We declare under our sole responsibility that /**

the medical device: / Dental Irrigation Needle  
specification: / Flexi Irrigation Needle, Metal Irrigation Needle

classification and Rules: / Class I – Rule 5 subclause 1  
according to Annex VIII of the Regulation (EU) 2017/745  
EMDN code / description: / A010602 - DENTAL IRRIGATION NEEDLES  
Basic -UDI-DI 69705149109ITC5

**meets the regulatory requirement of the European Union Medical Device Regulation (EU)2017/745**

Conformity assessment procedure: / Annex II+III of Regulation (EU) 2017/745  
Applied standards: / EN ISO13485:2016+AC:2018+A11:2021, EN ISO 14971:2019, EN ISO 15223-1:2021, EN ISO 20417:2021

Authorised Representative: /



**Share Info GmbH**

**Address: Am Schulzentrum 12, 41564 Kaarst, Germany**  
**SRN: DE-AR-000005132**

Signature: \_\_\_\_\_ Date: 2025-4-16

Position: GM

Place: Shenzhen /China