

E.D.T.A
Canal Irrigasyon and Cleaning Solution



INSTRUCTIONS FOR USE



Description

A cleaning solution containing 5% Ethylene diamine tetraacetic acid. Used in canal irrigation and disinfection.

Composition

Ethylene diamine tetraacetic acid, Sodium hydroxide, Sodium salt, Distilled water

Indications

Used in In root canal treatment, in cleaned by washing of the root canal, in humidification of the root canal.
Soft and hard tissue chemically and physically removes residues.

Application

After the isolation Rubberdam entered into the root canal system.
Pulp chamber is evacuated by an excavator and washed with irrigation solutions.
Irrigation clinically applied with side disposable ferfore needles.
Facilitated entry into the root canal using dental needle.
The needle tip is placed in the root canal drained with moderate pressure solution.
Overflowing portion is impregnated with an aspirator.
Canal file is placed in E.D.T.A full canal is continued cleaning and shaping process.
This process continues until the last use of the instrument.

Side effects

There are no known side effects

Storage

Protect from sunlight Store at 1–30 °C / 33,8–86 °F . Keep in cool climates.

Cautions

Keep material out of children's reach!
For use in dentistry only!
Irritant be; Avoid contact with skin, eyes and mucous membranes.
Certainly drinking.

Disposal: Disposal must be made according to official regulations. Take to an approved landfill or a waste incineration plant, under conditions approved by the local authority.


Uncleaned packagings: Disposal must be made according to official regulations

Processing should be carried out strictly according to the Instructions for Use. Liability cannot be accepted for damage resulting from failure to observe the Instructions or the stipulated area of application. The user is responsible for testing the material for its suitability and use for any purpose not explicitly stated in the Instructions. Descriptions and data constitute no warranty of attributes.

Expiry Date

From the moment the cover is opened should be used within 6 months.
Manufacturing dates and LOT number on the top of the product cover.

Date instructions for use prepared: Rev 04 TD-02-IFU-06 21.03.2019 Ref. 0909



IMICRYL FETİH MH. MAHIR SK. NO:5/201 TEL. +90 444 71 30
42030 KONYA TURKEY www.imicryl.com

E.D.T.A
Canal Irrigasyon and Cleaning Solution



INSTRUCTIONS FOR USE



Description

A cleaning solution containing 5% Ethylene diamine tetraacetic acid. Used in canal irrigation and disinfection.

Composition

Ethylene diamine tetraacetic acid, Sodium hydroxide, Sodium salt, Distilled water

Indications

Used in In root canal treatment, in cleaned by washing of the root canal, in humidification of the root canal.
Soft and hard tissue chemically and physically removes residues.

Application

After the isolation Rubberdam entered into the root canal system.
Pulp chamber is evacuated by an excavator and washed with irrigation solutions.
Irrigation clinically applied with side disposable ferfore needles.
Facilitated entry into the root canal using dental needle.
The needle tip is placed in the root canal drained with moderate pressure solution.
Overflowing portion is impregnated with an aspirator.
Canal file is placed in E.D.T.A full canal is continued cleaning and shaping process.
This process continues until the last use of the instrument.

Side effects

There are no known side effects

Storage

Protect from sunlight Store at 1–30 °C / 33,8–86 °F . Keep in cool climates.

Cautions

Keep material out of children's reach!
For use in dentistry only!
Irritant be; Avoid contact with skin, eyes and mucous membranes.
Certainly drinking.

Disposal: Disposal must be made according to official regulations. Take to an approved landfill or a waste incineration plant, under conditions approved by the local authority.


Uncleaned packagings: Disposal must be made according to official regulations

Processing should be carried out strictly according to the Instructions for Use. Liability cannot be accepted for damage resulting from failure to observe the Instructions or the stipulated area of application. The user is responsible for testing the material for its suitability and use for any purpose not explicitly stated in the Instructions. Descriptions and data constitute no warranty of attributes.

Expiry Date

From the moment the cover is opened should be used within 6 months.
Manufacturing dates and LOT number on the top of the product cover.

Date instructions for use prepared: Rev 01 03.07.2020 TD-02-IFU-06 Ref No:0909



IMICRYL FETİH MH. MAHIR SK. NO:5/201 TEL. +90 444 71 30
42030 KONYA TURKEY www.imicryl.com